



EU Quality Management System Certificate (MDR)

Pursuant to Regulation (EU) 2017/745 on Medical Devices, Annex IX Chapters I and III
(Class IIa and Class IIb Devices)

No. G10 078672 0022 Rev. 00

Manufacturer:

Shenzhen Urion Technology Co., Ltd.

Floor 4-6th of Building D
Jiale Science & Technology Industrial Zone
No.3, ChuangWei Road
Heshuikou Community, MaTian Street
GuangMing New District
518106 Shenzhen
PEOPLE'S REPUBLIC OF CHINA

SRN Manufacturer - CN-MF-000018740

Authorized Representative:

Wellkang Ltd
Enterprise Hub, NW Business Complex, 1 Beraghmore Road,
Derry, BT48 8SE, UNITED KINGDOM - NORTHERN IRELAND

The Certification Body of TÜV SÜD Product Service GmbH certifies that the manufacturer has established, documented and implemented a quality management system as described in Article 10 (9) of the Regulation (EU) 2017/745 on medical devices. Details on device categories covered by the quality management system are described on the following page(s). The Report referenced below summarises the result of the assessment and includes reference to relevant CS, harmonized standards and test reports. The conformity assessment has been carried out according to Annex IX Chapter I and III of this regulation with a positive result.

The quality management system assessment was accompanied by the assessment of technical documentation for devices selected on a representative basis.

The certified quality management system is subject to periodical surveillance by TÜV SÜD Product Service GmbH. The surveillance assessment shall also include an assessment of the technical documentation for the device or devices concerned on the basis of further representative samples. All applicable requirements of the Testing, Certification, Validation and Verification Regulations TÜV SÜD Group have to be complied with.

For details and certificate validity see: www.tuvsud.com/ps-cert?q=cert:G10 078672 0022 Rev. 00

Report No.: GZ2217903

Valid from: 2024-03-01

Valid until: 2029-02-28

Christoph Dicks
Head of Certification/Notified Body

Issue date: 2024-03-01



Benannt durch/Designated by
 Zentralstelle der Länder
 für Gesundheitsschutz
 bei Arzneimitteln und
 Medizinprodukten
 www.zlg.de
 BS-MDR-099



Product Service

EU Quality Management System Certificate (MDR)

Pursuant to Regulation (EU) 2017/745 on Medical Devices, Annex IX Chapters I and III
 (Class IIa and Class IIb Devices)

No. G10 078672 0022 Rev. 00

Classification: Class IIa
Device Group: V03010102 - DIGITAL THERMOMETERS
Intended Purpose: -

Classification: Class IIa
Device Group: Z12030205 - NON-INVASIVE BLOOD PRESSURE GAUGES
Intended Purpose: -

The validity of this certificate depends on conditions and/or is limited to the following: -

Revision History:

Rev.	Dated	Report	Description
00	2024-03-01	GZ2217903	Initial issuance